



International Journal of Multidisciplinary Research and Growth Evaluation.

Method development and validation for the quantitative estimation of Trametinib in API form and marketed tablet dosage form by RP-HPLC

K Deepthi ^{1*}, Bommagouni Ramya ², Jangili Priyanka ³, Lingampally Pranali ⁴, Thalla Harisha ⁵, Ganta Brundharika ⁶

¹ Department of Pharmaceutics, Bojjam Narasimhulu Pharmacy College for Women, Sayeedabad, Hyderabad, Telangana, India

²⁻⁶ Bojjam Narasimhulu Pharmacy College for Women, Road Number 2, Vinayak Nagar, Vani Nagar, Saroor Nagar West, Sayeedabad, Hyderabad, Telangana, India

* Corresponding Author: **K Deepthi**

Article Info

ISSN (online): 2582-7138

Volume: 04

Issue: 01

January-February 2023

Received: 03-01-2023;

Accepted: 21-01-2023

Page No: 369-374

DOI:

<https://doi.org/10.54660/IJMRGE.2023.4.1.369-374>

Abstract

The development of analytical methods is in need for the estimation of Trametinib in pure and different pharmaceutical formulations. A simple, sensitive, rapid, accurate, precise and economic chromatographic method was developed and validated for Trametinib in pure and pharmaceutical formulations. The proposed method was validated according to the International Conference on Harmonization (ICH) guidelines. This separation is performed on Symmetry ODS (C18) RP Column, 250 mm x 4.6 mm, 5 μ m Column and mobile phase consists of ACN: Methanol: 0.1% OPA in the ratio of 60:30:10v/v/v at flow rate of 1.0ml/min. The wave length (λ_{max}) used for the estimation of Trametinib is 267 nm by chromatographic method. The linearity of the calibration curve was validated by the high values of the correlation coefficient of regression. The percentage of Trametinib recovered was found to be within the limits i.e. 98-102% for Trametinib. LOD and LOQ values for Trametinib were found to be 0.08 μ g/ml and 0.24 μ g/ml respectively. The developed methods are simple and suitable for the determination of Trametinib in pure and pharmaceutical preparations.

Keywords: Trametinib, RP-HPLC, ICH guidelines, robustness, linearity, precision

Introduction

Trametinib is an orally bioavailable inhibitor of mitogen-activated protein kinase (MAP2K; MAPK/ERK kinase; MEK) 1 and 2, with potential antineoplastic activity. Upon oral administration, Trametinib ^[1] specifically binds to and inhibits MEK 1 and 2, resulting in an inhibition of growth factor-mediated cell signaling and cellular proliferation in various cancers. MEK 1 and 2, dual specificity serine/threonine and tyrosine kinases often upregulated in various cancer cell types, play a key role in the activation of the RAS/RAF/MEK/ERK signaling pathway that regulates cell growth. Trametinib is an anticancer agent which causes apoptosis (or programmed cell death) and inhibits cell proliferation, which are both important in the treatment of malignancies. Trametinib ^[2] is a reversible, allosteric inhibitor of mitogen-activated extracellular signal regulated kinase 1 (MEK1) and MEK2 activation and of MEK1 and MEK2 kinase activity. MEK proteins are upstream regulators of the extracellular signal-related kinase (ERK) pathway, which promotes cellular proliferation. Trametinib helps with melanoma with the BRAF V600E or V600K as the mutation results in the constitutive activation of the BRAF pathway which includes MEK1 and MEK2. Trametinib is indicated for the treatment of unresectable or metastatic melanoma with BRAF V600E or V600K mutations, as detected by an FDA-approved test [FDA]. Trametinib ^[3] is used alone or in combination with Dabrafenib (Tafinlar) to treat a certain types of melanoma (a type of skin cancer) that cannot be treated with surgery or that has spread to other parts of the body. The IUPAC Name of Trametinib is N-[3-[3-cyclo propyl-5-(2-fluoro-4-iodo anilino)-6, 8-dimethyl-2, 4, 7-trioxo pyrido [4, 3-d] pyrimidin-1-yl] phenyl] acetamide. The Chemical Structure of Trametinib is as following growth opportunities, according to several scholars.




Fig 1: Chemical Structure of Trametinib

Table 1: List of Instrument used

S. No.	Instruments/Equipments/Apparatus
1.	HPLC with Empower2 Software with Isocratic with UV-Visible Detector (Waters).
2.	T60-LAB INDIA UV – Vis spectrophotometer
3.	Electronic Balance (SHIMADZU ATY224)
4.	Ultra Sonicator (Wensar wuc-2L)
5.	Thermal Oven
6.	Symmetry ODS RP C ₁₈ , 5µm, 15mm x 4.6mm i.d.
7.	pH Analyzer (ELICO)
8.	Vacuum filtration kit (BOROSIL)

Materials and Methods

Instruments used

Chemicals / Reagents Used

Table 2: List of Chemicals used

S.No.	Name	Specifications		Manufacturer/Supplier
		Purity	Grade	
1.	Doubled distilled water	99.9%	HPLC	Sd fine-Chem ltd; Mumbai
2.	Methanol	99.9%	HPLC	Loba Chem; Mumbai.
3.	Dipotassium hydrogen orthophosphate	96%	A.R.	Sd fine-Chem ltd; Mumbai
4.	Acetonitrile	99.9%	HPLC	Loba Chem; Mumbai.
5.	Potassium dihydrogen orthophosphate	99.9%	A.R.	Sd fine-Chem ltd; Mumbai
6.	Sodium hydroxide	99.9%	A.R.	Sd fine-Chem ltd; Mumbai
7.	Hydrochloric acid	99.9%	A.R.	Loba Chem; Mumbai.
8.	Hydrogen Peroxide	99.9%	A.R.	Loba Chem; Mumbai.

Method Development and its Validation for Trametinib by RP-HPLC

Method Development

Selection of Wavelength

The standard & sample stock solutions were prepared separately by dissolving standard & sample in a solvent in mobile phase diluting with the same solvent.(After optimization of all conditions) for UV analysis. It scanned in the UV spectrum in the range of 200 to 400nm. This has been performed to know the maxima of Trametinib, so that the same wave number can be utilized in HPLC UV detector for estimating the Trametinib. The scanned UV spectrum is attached in the following page.

Sample & Standard Preparation for the UV-Spectrophotometer Analysis

25 mg of Trametinib standard was transferred into 25 ml volumetric flask, dissolved & make up to volume with mobile phase. Further dilution^[4] was done by transferring 0.5 ml of the above solution into a 10ml volumetric flask and make up to volume with mobile phase.




Fig 2: UV Spectrum for Trametinib

Observation: While scanning the Trametinib solution we observed the maxima at 235nm. The UV spectrum has been recorded on T60-LAB INDIA make UV – Vis spectrophotometer model UV-2450.

Optimization of Chromatographic Conditions

The chromatographic conditions^[5] were optimized by different means. (Using different column, different mobile phase, different flow rate, different detection wavelength & different diluents for sample preparation^[6] etc.

Table 3: Summary of Process Optimization

Column Used	Mobile Phase	Flow Rate	Wave length	Observation	Result
Symmetry C ₁₈ , ODS, Reverse Phase, 250 mm x 4.6 mm, 5µm, Column.	Methanol: Acetonitrile = 40: 60	1.0ml/min	235nm	Very Low response	Method rejected
Symmetry C ₁₈ , ODS, Reverse Phase, 250 mm x 4.6 mm, 5µm, Column.	Methanol: Acetonitrile = 55: 45	1.0ml/min	235nm	Low response	Method rejected
Symmetry C ₁₈ , ODS, Reverse Phase, 250 mm x 4.6 mm, 5µm, Column.	Acetonitrile: Water = 50:50	1.0ml/min	235nm	Tailing peaks	Method rejected
Symmetry C ₁₈ , ODS, Reverse Phase, 250 mm x 4.6 mm, 5µm, Column.	Methanol: Water = 70:30	1.0ml/min	235nm	Resolution was not good	Method rejected
Symmetry C ₁₈ , ODS, Reverse Phase, 250 mm x 4.6 mm, 5µm, Column.	ACN: Methanol: 0.1% OPA = 70:25:5	1.0ml/min	235nm	Tailing peak	Method rejected
Symmetry C ₁₈ , ODS, Reverse Phase, 250 mm x 4.6 mm, 5µm, Column.	ACN: Methanol: 0.1% OPA = 60:30:10	1.0ml/min	235nm	Nice peak	Method accepted

Summary of optimized chromatographic conditions

The Optimum Chromatographic conditions ^[7] obtained from

experiments can be summarized as below:

Table 4: Summary of Optimised Chromatographic Conditions

Mobile phase	ACN: Methanol: 0.1% OPA = 60:30:10
Column	Symmetry ODS (C ₁₈) RP Column, 250 mm x 4.6 mm, 5µm
Column Temperature	Ambient
Detection Wavelength	235 nm
Flow rate	1.0 ml/ min.
Run time	06 min.
Temperature of Auto sampler	Ambient
Diluent	Mobile Phase
Injection Volume	10µl
Type of Elution	Isocratic
Retention time	2.570 minutes




Fig 3: Chromatogram of Trametinib in Optimized Condition

Observation: The selected and optimized mobile phase ^[8] was ACN: Methanol: 0.1% OPA = 60:30:10 and conditions optimized were flow rate (1.0 ml/minute), wavelength (235nm), Run time was 06 mins. Here the peaks were separated and showed better resolution, theoretical plate count and symmetry ^[9]. The proposed chromatographic conditions were found appropriate for the quantitative determination of the drug.

Preparation of Mobile Phase

600ml of HPLC Grade Acetonitrile, 300ml of HPLC Grade Methanol and 100ml 0.1% OPA were mixed well and degassed in ultrasonic water bath for 15 minutes. The solution was filtered through 0.45 µm filter under vacuum filtration ^[10].

Validation of Method

1. Accuracy

Recovery study

To decide the exactness of the proposed strategy, recuperation thinks about were done by including diverse sums (80%, 100%, and 120%) of unadulterated medication of TRAMETINIB were taken and added to the pre-broke down plan of fixation 10µg/ml. From that rate recuperation ^[11] esteems were computed. The outcomes were appeared in table-5.

Table 5: Readings of Accuracy

Conc. In ppm	Conc. Found	Peak Area	% Recovery
8	8.035	161523	100.437
8	8.153	163815	101.912
8	8.061	162023	100.762
		Avg.	101.037
		S.D	0.775
		%RSD	0.767046
Conc. In ppm	Conc. Found	Peak Area	% Recovery
10	9.930	198315	99.30
10	10.033	200320	100.33
10	10.044	200540	100.44
		Avg.	100.0233
		S.D	0.628835
		%RSD	0.628688
Conc. In ppm	Conc. Found	Peak Area	% Recovery
12	11.981	238151	99.841
12	12.066	239819	100.55
12	12.215	242712	101.791
		Avg.	100.7273
		S.D	0.987021
		%RSD	0.979894

2. Precision

2.1. Repeatability

The precision ^[12] of each method was ascertained separately from the peak areas & retention times obtained by actual

Determination of six replicates of a fixed amount of drug, Trametinib (API). The percent relative standard deviation ^[13] was calculated for Trametinib are presented in the table-6.

Table 6: Readings of Repeatability

HPLC Injection Replicates of Trametinib	Retention Time (Minutes)	Peak Area (AUC)
Replicate – 1	2.572	197236
Replicate – 2	2.570	197762
Replicate – 3	2.573	195969
Replicate – 4	2.570	194724
Replicate – 5	2.574	198327
Replicate – 6	2.573	198711
Average		197121.5
Standard Deviation		1515.213
% RSD		0.768667

2.2. Intermediate Precision

2.2.1. Intra-assay & inter-assay

The intra & inter day variation ^[14] of the method was carried out & the high values of mean assay ^[15] & low values of standard deviation & % RSD (% RSD < 2%) within a day & day to day variations for Trametinib revealed that the proposed method is precise.

Table 7: Results of Intra-Assay & Inter-Assay

Conc. of Trametinib(API) (µg/ml)	Observed Conc. of Trametinib (µg/ml) by the proposed method			
	Intra-Day		Inter-Day	
	Mean (n=6)	% RSD	Mean (n=6)	% RSD
8	7.46	0.62	8.05	0.96
10	10.87	0.85	9.43	0.71
12	11.81	0.92	12.04	0.65

3. Linearity & Range:

The calibration curve showed good linearity ^[16] in the range of 6 – 14 µg/ml, for Trametinib (API) with correlation coefficient ^[17] (r^2) of 0.999 (Fig-4). A typical calibration curve ^[18] has the regression equation of $y = 19423x + 5444$ for Trametinib.




Fig 4: Calibration Curve of Trametinib (API).

Table 8: Linearity Results

CONC.(µg/ml)	MEAN AUC (n=6)
0ppm	0
6ppm	129013
8ppm	166523
10ppm	198315
12ppm	234151
14ppm	275819

4. Method Robustness: Influence of small changes in chromatographic conditions such as change in flow rate (± 0.1 ml/min), Wavelength of detection (± 2 nm) & organic phase in mobile phase ($\pm 5\%$) studied to determine the robustness ^[19] of the method are also in favour of (Table-9, % RSD < 2%) the developed RP-HPLC method for the analysis of Trametinib (API).

Table 9: Result of Method Robustness Test

Change in parameter	% RSD
Flow (1.1 ml/min)	0.68
Flow (0.9 ml/min)	0.39
More Organic	0.54
Less Organic	0.63
Wavelength of Detection (237 nm)	0.91
Wavelength of detection (233 nm)	0.93

5. LOD & LOQ

The Minimum concentration level at which the analyte can be reliable detected (LOD ^[20]) & quantified (LOQ ^[21]) were found to be 0.08 & 0.24µg/ml respectively.

6. System Suitability Parameter: Framework appropriateness testing ^[22-26] is an essential piece of numerous scientific techniques. The tests depend on the idea that the gear, hardware, explanatory activities and tests to be broke down establish a vital framework that can be assessed all things considered. Following framework appropriateness test parameters were built up. The information is appeared in Table-10.

Table 10: Data of System Suitability Parameter

S. No.	Parameter	Limit	Result
1	Resolution	$R_s > 2$	8.47
2	Asymmetry	$T \leq 2$	Trametinib=0.23
3	Theoretical plate	$N > 2000$	Trametinib=2987
4	Tailing Factor	$T < 2$	Trametinib=1.17

7. Estimation of Trametinib in Pharmaceutical Dosage Form

Twenty pharmaceutical dosage forms ^[27] were taken and the I.P. strategy was taken after to decide the normal weight. Above measured tablets were at last powdered and triturated well. An amount of powder proportionate to 25 mg of medications were exchanged to 25 ml volumetric flagon, make and arrangement was sonicated for 15 minutes, there after volume was made up to 25 ml with same dissolvable. At that point 10 ml of the above arrangement was weakened to 100 ml with versatile stage. The arrangement was separated through a layer channel (0.45 µm) and sonicated to degas. The arrangement arranged was infused in five reproduces into the HPLC framework and the perceptions were recorded. A copy infusion of the standard arrangement was additionally infused into the HPLC framework ^[28] and the peak regions were recorded. The information is appeared in Table-11.

$$\text{Assay \%} = \frac{\text{AT} \quad \text{WS} \quad \text{DT} \quad \text{P}}{\text{AS} \quad \text{DS} \quad \text{WT} \quad 100} \times \text{Avg. Wt} = \text{mg/tab}$$

Where:

AT = Peak Area of medication acquired with test arrangement

AS = Peak Area of medication acquired with standard ^[29] arrangement

WS = Weight of working standard taken in mg

WT = Weight of test taken in mg

DS = Dilution of Standard arrangement

DT = Dilution of test arrangement

P = Percentage virtue of working standard

Table 11: Recovery Data for estimation Trametinib in Mekinist

Brand Name of Trametinib	Labelled amount of Drug (mg)	Mean (\pm SD) amount (mg) found by the proposed method (n=6)	Assay % (\pm SD)
Mekinist Tablets (GlaxoSmithKline)	2mg	1.885 (\pm 0.875)	99.89 (\pm 0.452)

Result & Discussion: The amount of drug in Mekinist Tablets was found to be 1.885 (\pm 0.875) mg/tab for Trametinib & % assay ^[30-31] was 99.89 %.

Summary and Conclusion

To develop a precise, linear, specific & suitable stability indicating RP-HPLC method for analysis of Trametinib, different chromatographic conditions were applied & the results observed are presented in previous chapters. Isocratic elution is simple, requires only one pump & flat baseline separation for easy and reproducible results. So, it was preferred for the current study over gradient elution. In case of RP-HPLC various columns are available, but here Symmetry ODS RP C₁₈, 5 μ m, 15mmx4.6mm i.d. Column was preferred because using this column peak shape, resolution and absorbance were good.

Discovery wavelength was chosen in the wake of examining the standard arrangement of medication more than 200 to 400nm. From the U.V range of Trametinib it is apparent that a large portion of the HPLC works can be proficient in the wavelength scope of 210-300 nm helpfully. Further, a stream rate of 1 ml/min and an infusion volume of 10 μ l were observed to be the best investigation. The outcome demonstrates the created technique is amazingly, one more reasonable strategy for measure and dependability related debasement examines which can help in the investigation of Trametinib in various details.

References

1. <https://go.drugbank.com/drugs/DB08911>
2. <https://pubchem.ncbi.nlm.nih.gov/compound/Trametinib>
3. <https://en.wikipedia.org/wiki/Trametinib>
4. Morgan DJ. Fraction collector (post on Flickr). Flickr [Internet]. 2015 Oct 28 [cited 2024 Jun 28]. Available from: <https://www.flickr.com/photos/davidjmorgan/21654734513>.
5. Karger BL. HPLC: Early and recent perspectives. Journal of Chemical Education. 1997;74:45.
6. Henry RA. The early days of HPLC at Dupont. Chromatography Online. Avanstar Communications Inc.; 2009.
7. Iler RK. The Chemistry of Silica. New York: John Wiley

- & Sons; 1979.
8. Karger BL, Berry LV. Rapid liquid-chromatographic separation of steroids on columns heavily loaded with stationary phase. Clinical Chemistry. 1971;17(8):757-64.
9. Giddings JC. Dynamics of Chromatography, Part I. Principles and Theory. New York: Marcel Dekker, Inc.; 1965;281.
10. Etre C. Milestones in chromatography: The birth of partition chromatography [PDF]. LCGC. 2016;19(5):506-12.
11. Martin AJP, Synge RLM. Separation of the higher monoamino-acids by counter-current liquid-liquid extraction: the amino-acid composition of wool. Biochemical Journal. 1941;35(1-2):91-121.
12. Lindsay S, Kealey D. High Performance Liquid Chromatography. Wiley. In: Hung LB, Parcher JF, Shores JC, Ward EH. Theoretical and experimental foundation for surface-coverage programming in gas-solid chromatography with an adsorbable carrier gas. Journal of the American Chemical Society. 1987;110(11):1090-6.
13. Journal of Pharmaceutical and Biomedical Analysis. 1999 Nov 1;21(2):371-82.
14. Tropical Journal of Pharmaceutical Research. 2009;8(5):449-54.
15. Sankar R. Instrumental Method of Analysis. 3rd ed. New Delhi: Pharmaceutical Press; 2017;18(3):18-6.
16. Snyder LR. Practical HPLC Method Development. 2nd ed. New York: Wiley-Interscience; 1997;503.
17. U.S. Department of Health and Human Services FDA. Guidance for Industry: Analytical Procedure and Method Validation. Rockville: FDA; 2000.
18. ICH. Validation of Analytical Procedures: Methodology. ICH Harmonized Tripartite Guideline Q2B; 1996.
19. Sharma BK. Instrumental Methods of Chemical Analysis. Meerut: Goel Publishing House. 1997;75(8):113-5.
20. Ewing GW. Instrumental Methods of Chemical Analysis. 5th ed. New York: McGraw-Hill; 1985;1.
21. Higuchi T, Brochmann-Hanffen E, Hanssen H. Pharmaceutical Analysis. 1st ed. New York: Wiley; 1961;1-10.
22. Beckett AH, Stenlake JB. Practical Pharmaceutical Chemistry. 4th ed. Vol. 2. London: Athlone Press; 1988;275-98.
23. Inforum. Quality Assurance: Worth the Effort. 2003;7(4).
24. Sethi PD. Quantitative Analysis of Drugs in Pharmaceutical Formulation. 3rd ed. New Delhi: CBS Publishers. 1997;1(21):51-6.
25. ICH. Text on Validation of Analytical Procedures. ICH Harmonized Tripartite Guidelines; 1994.
26. ICH. Validation of Analytical Procedures: Methodology. ICH Guidelines Q2B. 1996;11(CPMP/ICH/281/95).
27. Gupta V, Rajput S, Sharma A, Khare P. Development and validation of HPLC method - A review. International Research Journal of Pharmaceutical and Applied Sciences. 2012;2(4):17-25.
28. Bhardwaj SK, Dwivedi J, Mehta A, Kurmi M. A review: HPLC method development and validation. International Journal of Analytical and Bioanalytical Chemistry. 2015;5(3):1-10.
29. Chromacademy. Method Development: A Guide to

- Basics Quantitative & Qualitative HPLC, LC, GC.
30. Sonawane LV. Bioanalytical method validation and its pharmaceutical application - A review. *Pharmaceutica Analytica Acta*. 2014;5(3):1-5.
 31. ICH. Validation of Analytical Procedures: Text and Methodology. ICH Topic Q 2 (R1).
 32. Ramakrishna B, Babu NS, Naidu NV. Method development and validation of RP-HPLC for assay of Trametinib in pharmaceutical dosage form. *European Journal of Pharmaceutical and Medical Research*. 2018;5(3):318-25.
 33. Illendula S, Therisa GM. Simultaneous estimation of latest analytical method improvement and validation of Dabrafenib and Trametinib by means of high-performance liquid chromatography. *World Journal of Pharmacy and Pharmaceutical Sciences*. 2021;10(3):1978-2004.